

based on Tipni  
date - 21-6-2022

SGH/SUR/MJPJAY/UM62/2022


Date - 21-6-2022


Sassoon Hospital, Pune  
quotation submitted at  
Administrative office


Last date - 30-6-2022  
5:00 PM

Uncemented Total Hip Replacement -Metal on poly

Cemented Total Hip Replacement-Metal on poly			
S.NO	Component	Specification	Material
1	Acetaubular Cup	Cemented All polyethylene XLPE Acetabular Cup	XLPE poly ethylene liner
2	Femoral stem	Titanium/Cobalt chrome cemented Triple -taper design or CPT design	Cobalt chrome CoCr/Titanium
3	Head	SST/Cobalt chrome head (28/32)	Cobalt chorome CoCr
4	Cement	1 packets of high viscosity bone cement and 1 packet of low viscosity bone cement	
5	Others	Saw blade-1	
All Implants should be US FDA approved			
Complete Instrument set should be provided at the time of surgery			
Cordless battery operated hand drill +saw+saw blade			
Complete range of implant sizes should be provided at the time of surgery.			

  
Associate Professor  
Orthopaedic Dept.  
B.J.M.C. & S.G.H. Pune

  
सहाय्यक प्राध्यापक  
अस्थिव्यंगोपचार विभाग  
\* श्री. शास. महाविद्यालय, पुणे

  
Associate Professor  
Orthopaedic Dept.  
B.J.M.C. & S.G.H. Pune