based on TSPM
date - 21-6-2022

SGH/SUR/MJPJAY/4463/2022 Date - 21-6-2022

Last date - 30-6-2022
5:00 PM

Sassoon Hospital, Pune
quotation submitted
at Administrative office

Uncemented Total Hip Replacement –Metal on poly

1	Acetabular Cup	Uncemented beaded porous coated acetabular cup with three screw holes Highly Polisehd Inner Surface	Titanium
2	Femoral stem	Uncemented stem with lateral offset option, 12/14 taper and low profile lateral shoulder (Fully HA coated) and option for Std and high offset	Titanium
3	Acetabular Liner	Highly crossed linked acetabular liner with micrastable locking mechanism	XLPE polyethylene liner
4	Head	SST/Cobalt chrome head (28/32/36)	Cobalt chrome COCR
		All Implants should be US FDA approved	
		Complete Instrument set should be provided at the time of surgery	
		Cordless battery operated hand drill +saw+saw blade	
		Complete range of implant sizes should be provided at the time of surgery.	

Associate Professor
Orthopaedic Dept.
B.J.M.C. & S.G.H. Pune

Associate Professor
Orthopaedic Dept.
B.J.M.C. & S.G.H. Pune

सहाय्यक प्राध्यापक
अस्थिव्यंगोपचार विभाग
बै.जी. शास. महाविद्यालय, पुणे