



Sassoon General Hospital, Pune

Jayprakash Narayan Road, Pune Railway Station Near, Pune 411 001

Purchase Section

Tel No.020-26102306 Email. www.bjmc-pune.org/www.sassoonpurchase@gmail.com

Fax No.020-26053452

No.SGH/PC/Paed Dept/Hot air oven/Quotation/945 /18

Dt.13 /12/2018

Online Quotation

Subject : Quotation for Supply Hot Air Oven Paediatrics Dept.

This Hospital has to purchase Hot Air Oven Machines in Paediatrics Dept. mentioned below. Therefore, please send your quotation in sealed envelope before the last date in the name of Dean, Sassoon General Hospital, Pune. The terms and conditions are as under. Quotation Letter the Mention CST/MST Vat/Tin No, & Pan No.

Last Date to Submit Quotation : Dt. 21 /12/2018

Sr.No.	Description	Per Qty. Rate
01	Hot Air Oven	01
	<p>Specifications</p> <p>1) Cabinet – Double walled cabinet with outer wall made of mild steel, and inner chamber can be of mild steel, or stainless steel, as required-</p> <p>2)Insulation – Glass wool fiked between the walls to a minimum thickness of 2 inch.</p> <p>3)Door – The door is fitted on heavy chrome plated brass hinges, and it contains a spring and roller type latch</p> <p>4)Finish – Outer side is painted with enamel paint. The inner side (in case of mild steel) is painted with heat-resistive aluminum paint.</p> <p>5)Heating – heating is carried out by specially designed low-rating heaters.</p> <p>6)Control – The temperature is controlled by a specially designed low-rating heaters.</p> <p>7)Temp. Range – For PEW-180 : 5' C above the ambient temperature to 180' C Max. For PEW-205:5'C above the ambient temperature to 250' C max. (up to 300' C can be provided, at an extra cost)</p> <p>8)Accuracy – Better than +2' C below 80' C and better than +1'C above 80'C</p>	–

Terms & Conditions :-

- 1) Rights to Accept or reject quotation are reserved by this office.
- 2) Quotation received after last date will not be considered.
- 3) Mention Taxes and other charges if any separately.
- 4) The Sale tax registration number is to be quoted in quotation otherwise your quotation will not be considered.
- 5) Demonstration of machine is compulsory.
- 6) Quantity is flexible.
- 7) Warrantee – One year from date of installation.

Y. S. S.
DEAN

Sassoon General Hospital, Pune