



Sassoon General Hospital, Pune

Jayprakash Narayan Road, Pune Railway Station Near, Pune 411 001

Purchase Section

Tel No.020-26102306 Email. www.bjmcpcune.org/www.sassoonpurchase@gmail.com

Fax No.020-26053452

No.SGH/PS/Infrared Thermometer/Quotation/5447/20

Dt. 6/07/2020

Online Quotation

Subject : Quotation for Infrared Thermometer

This Hospital has to purchase the Infrared Thermometer corona virus Dept. mentioned below. Therefore, please send your quotation in sealed envelope before the last date in the name of Dean, Sassoon General Hospital, Pune. The terms and conditions are as under. Mention GST, Vat/Tin No, & Pan No. in Quotation letter.

Last Date to Submit Quotation : Dt.

06/2020 14-7-2020

Sr.No.	Name of Machine With Attach Specification	Per Qty. Rate
01	Infrared Thermometer	01

Attached Specification

Terms & Conditions :-

- 1) Rights to Accept or reject quotation are reserved by this office.
- 2) Quotation received after last date will not be considered.
- 3) Mention Taxes and other charges if any separately.
- 4) The GST registration number is to be quoted in quotation otherwise your quotation will not be considered.
- 5) Demonstration of machine is compulsory.
- 6) Quantity is flexible.
- 7) Warrantee – One year from date of installation.

DEAN


Sassoon General Hospital, Pune

Specification

Equipment: Infrared Thermometer

Technical Specification:

1. Touch-Free Infrared Forehead Thermometer measures body and object and liquid temperatures.
2. Safe, sanitary and noninvasive temperature readings - body temperature measurements taken at forehead.
3. Fever icon indicator can quickly determine normal or high measurements. Illuminated, backlit display for low light use.
4. CE, FCC, RoHS Certification
5. Built-in laser pointer, precise non-contact infrared measurement.
6. Calibration (default 0), you can set when there is a temperature deviation;
7. Memories
8. Measurement modes selection for various needs conveniently : surface temperature (Room, milk, water) & Body Temperature
9. 2 temperature units for conversion. °C and °F Switchable.
10. Touch free, 0.5 second get the measuring result
11. Alarm Value Setting, Selection for buzzer


Professor
Dept. of Anaesthesiology
B.J. Medical College &
Sassoon General Hospitals, Pune 411001